

Annexure-A

(Please refer to the Minutes of Meeting of Pre-bid meeting held on 20/10/2014)

Amendment No. 1

Subsequent to the Pre-Bid Meeting held on 20.10.2014 at 12:00 PM and the queries sought by prospective bidders, the following Amendments related – Tender Documents and Technical Specifications are being issued as under:


Tender Documents			
Sr. No.	Para / Sub Para Sr. No. Reference	As Provided	As amended
1.	Section VII 3 A Para 12	The system must be certified for clinical use of >25000 source transfers.	The source drive should be replaced Free of cost (FOC) up to ten years from date of installation.
2.	Section VII 3 B.1 Para 7	Should supply fast multi colour compatible Laser printer cum plotter to print out data and isodose curves	Should supply fast multi colour compatible Laser printer to print out data and isodose curves
3.	Section VII 3 B.2 Para 11	Treatment Planning System should be network with the Simulator, CT Scanner & MRI present in the institute.	Deleted
4.	Section VII 3 C Para 8	New Martinez Prostate applicators - 1 set.	Prostate applicators - 1 set.
5.	Section VII 4,Para 10	Appropriate Thimble Chamber with calibration certificate for Ir192 competeble with electrometer	Appropriate Thimble Chamber with calibration certificate for Ir192 competeble with above Supplied electrometer
6.	Section VII 4,Para 11	Digital pocket dosimeter	Digital pocket dosimeter 2 Nos
7.	Section VII 4,Para 14	Set of QA tool's required for brachytherapy commission including gafchromic films.	Set of QA tool's required for brachytherapy commissioning
8.	Section VII 6,Para 6.4	Comprehensive Training on application after installation for one month on site and at least 2 weeks off site or till the user is familiar with operation of the unit.	The treatment planning system and treatment unit's operational training should be provided to one Radiation Oncologist and one Medical Physicist at a center of repute for 1 week in the country of manufacturer and also on-site training of 1 week to staff of dept.
9.	Section VII 6,Para 6.5	Complete installation should include treatment of minimum 5 patients	Certificate of Acceptance shall be issued within 30 days from date of commissioning or minimum 1 fraction each on 5 different patients whichever comes earlier.

Besta
20/10/14

10.	Section X B Para iii	In case, bidder is a foreign manufacturer, it should have supplied and installed in last three years from the date of tender opening, at least 10 (four) similar equipment meeting major parameters of technical specifications, anywhere in Asia and which is functioning satisfactory.	In case, bidder is a foreign manufacturer, it should have supplied and installed in last three years from the date of tender opening, at least 10 (Ten) similar equipment meeting major parameters of technical specifications, anywhere in Asia and which is functioning satisfactory
-----	----------------------------	---	---

Note: Section VII, Technical Specifications are being repeated hence bidders are requested to read it appropriately.

Other terms and conditions of the tender document remains the same.
The bidders are requested to see the Health Website for corrigendum on <http://health.mp.gov.in> and also can be seen on website www.mpeproc.gov.in



Dr. B.S. Ohri)
Chief General Manager (Technical)
MPPHSCL, Bhopa