

Madhya Pradesh Public Health Services Corporation Limited
Bhopal

Ref.No:-06 /MPPHSCL/EQUIP/Laparoscope/2014

Date: 18.11.2014

Amendment No. 1

The Pre-Bid Meeting was held on 14.11.2014 at 12:00 PM and accordingly the following Amendments related to Technical Specifications are being issued:

Sr. No	Para / Sub Para Sr. No. Reference	As Provided	As Amended
1.	Section VII Technical Specification A Para 1	Telescope 0 degree with parallel eye piece, 10 to 12 mm diameter/straight Telescope Fiber optic light transmission incorporated; should be compatible with the commonly available light cables (necessary adaptors should be provided) Can be sterilized by autoclaving, cidex solutions and Formalin Chamber. Should have inbuilt 6 mm instrument channel for ring applicator as well as CO2 gas insufflations channel with stopcock, Working length of 270-275 mm	Telescope 0 degree with parallel eye piece, 10 to 12 mm diameter/straight Telescope Fiber optic light transmission incorporated; should be compatible with the commonly available light cables (necessary adaptors should be provided) Can be sterilized by autoclaving, cidex solutions and Formalin Chamber. Should have inbuilt 6 mm instrument channel for ring applicator, Working length of 270-275 mm
2.	Section VII Technical Specification B Para 1	Cold light source with dual control having 2 halogen lamps of 150 watts ratings each; operatable on 220-240V and 50 Hz supply. Backup light source (Mini Light Source) compatible with 6Volt battery and 100 Watts halogen lamp SP laparoscope is also to be provided. For both Cold light source and backup light source compatible LED lamps Would be preferred.	Cold light source with dual control having 2 Halogen lamps of 150 – 250 watts ratings each; operatable at 220-240V and on 50 Hz supply and Backup light source (mini light source) should be LED type, light weight, handy and rechargeable with 2 batteries and should be directly mountable on laparoscope/Telescope.
3.	Section VII Technical Specification B Para 2 UPS	Portable UPS having 6 Volt Lithium Battery matching with the mini light source	Deleted
4.	Section VII Technical Specification B add Para 6	-----	50 pieces of O-rings should be supplied as a start- up kit with each equipment.

Other terms and conditions of the tender document remains the same.

The bidders are requested to see the Health Website for corrigendum on <http://health.mp.gov.in> and also on www.mpeproc.gov.in



(Dr. Himanshu Jayswar)
General Manager (Procurement)
MPPHSCL, Bhopal