

Techno-Commercial Evaluation of Tender No. 15/MPPHSCL/ FP Kits /2015 dated 09.02.2015							
Corrigendum1 dated 25/02/2015,A23Corrigendum 2 dated 26/02/2015,Corrigendum 3 dated 27-02-15							
Bid Submission date & time: 28/02/2015 , 15:00 Hrs							
Bid opening date & time: 28/02/2015 , 15:30 Hrs							
MINILAP/IUCD/NSV kits with Accessories							
SCHEDULE	TECHNICAL SPECIFICATIONS FOR	Unit	Quantity	M/s Instrumed India International	M/s SISCO LATEX PVT LTD	M/S VIJAY SALES	M/S NEW VISION CORPORATION
SCHEDULE I	A) Minilap Kit (Set) :		1600 sets				
	1 Set consists of the following items:						
1	Sponge Holder 10 inch	Nos	2	Complies	Complies	Complies	Complies
2	Artery Forceps Straight 6 inch	Nos	4	Complies	Complies	Complies	Complies
3	Artery Forceps Curved 6 inch	Nos	2	Complies	Complies	Complies	Complies
4	Allis Forceps 6 inch	Nos	4	Complies	Complies	Complies	Complies
5	Babcock Forceps 6 inch	Nos	2	Complies	Complies	Complies	Complies
6	Right Angled retractor 6 inch	No	1	Complies	Complies	Complies	Complies
7	Right Angled retractor 8 inch	No	1	Complies	Complies	Complies	Complies
8	Tubal hook 8 inch	No	1	Complies	Complies	Complies	Complies
9	Needle Holder 8inch	No	1	Complies	Complies	Complies	Complies
10	Plain Forceps 6inch	No	1	Complies	Complies	Complies	Complies
11	Tooth Forceps 6inch	No	1	Complies	Complies	Complies	Complies
12	curved Scissor 8inch	No	1	Complies	Complies	Complies	Complies
13	Straight Scissor 8inch	No	1	Complies	Complies	Complies	Complies
14	Nel Cath 14fr-can be procured seperately	No	1	Complies	Complies	Complies	Complies
15	BP Handle No. 4	No	1	Complies	Complies	Complies	Complies
16	BP Handle Blade no. 22 -can be procured seperately	No	1	Complies	Complies	Complies	Complies
	DEMO		OK	OK	OK	OK	OK
SCHEDULE II	B) IUCD Kit (Set)		1600sets	OK	OK	OK	OK
	1 Set consists of the following items:						
1	Sponge Holder 10 inch	No	1	Complies	Complies	Complies	Complies
2	Sims speculum small size	No	1	Complies	Complies	Complies	Complies
3	Sims Speculum medium size	No	1	Complies	Complies	Complies	Complies
4	Allis Forceps 6 inch	Nos	2	Complies	Complies	Complies	Complies
5	Uterine Sound	No	1	Complies	Complies	Complies	Complies
6	Kellys Forceps 12 inch	No	1	Complies	Complies	Complies	Complies
7	Straight Scissor 6 inch	No	1	Complies	Complies	Complies	Complies
				OK	OK	OK	OK

SCHEDULE	TECHNICAL SPECIFICATIONS FOR	Unit	Quantity	M/s Instrumed India International	M/s SISCO LATEX PVT LTD	M/S VIJAY SALES	M/S NEW VISION CORPORATION
SCHEDULE III	C) NSV KIT (Set)						
	1 Set consists of the following items:		1600 sets	Complies	Complies	Complies	Complies
1	Dissecting Forceps	No	1	Complies	Complies	Complies	Complies
	Made of stainless steel, sharp tip, light weight.			Complies	Complies	Complies	Complies
	Approximate Dimension and weight are:			Complies	Complies	Complies	Complies
	Tip angle curvature : 40 degree on plane			Complies	Complies	Complies	Complies
	Inside Surface of blades: Smooth			Complies	Complies	Complies	Complies
	Length of blades (Jt.To tip) : 25 mm			Complies	Complies	Complies	Complies
	Length (Ratchet to tip) : 110 mm			Complies	Complies	Complies	Complies
	Ratchests : 2/3 ratchets			Complies	Complies	Complies	Complies
	Weight : 25-30 gms			Complies	Complies	Complies	Complies
2	Ringed Clamp :	No	1	Complies	Complies	Complies	Complies
	Made of stainless steel, and light weight.			Complies	Complies	Complies	Complies
	Approximate Dimension and weight are:			Complies	Complies	Complies	Complies
	Total Length: (Ratchet to tip) : 25 mm			Complies	Complies	Complies	Complies
	Inside dia of clamp ring : 4 mm			Complies	Complies	Complies	Complies
	Outside dia of rings : 6.5 mm			Complies	Complies	Complies	Complies
	Clamp ring slit distance : 0.5 mm			Complies	Complies	Complies	Complies
	Ratchets: 2/3 ratchets			Complies	Complies	Complies	Complies
3	Sponge Forceps 10 inch	No	2	Complies	Complies	Complies	Complies
4	Straight Scissor 6 inch	No	1	Complies	Complies	Complies	Complies
5	Artery forceps 6 inch	No	1	Complies	Complies	Complies	Complies
	DEMO			OK	OK	OK	OK
6	Essential Criteria						
	Surgical Instruments of the Kit should be CE Marked (Engraved on each instrument).			Only brand name engraved	CE marked & brand name engraved	CE marked & brand name engraved	CE marked & brand name engraved
	High Surgical grade stainless steel (350-450 grade)			OK	OK	OK	OK
	Company name and catalogue number should be engraved on each instrument.			OK	OK	OK	OK
	Physical demonstration of all instruments is mandatory.			OK	OK	OK	OK
	Manufacturer should be ISO certified.			ISO certificate	ISO certificate	ISO certificate	ISO certificate

SCHEDULE	TECHNICAL SPECIFICATIONS FOR	Unit	Quantity	M/s Instrumed India International	M/s SISCO LATEX PVT LTD	M/S VIJAY SALES	M/S NEW VISION CORPORATION
	OTHERS TERMS AND CONDITIONS						
	EMD			YES	YES	YES	YES
	Tender Form Section X			YES	YES	YES	YES
	Power of Attorney			Not Applicable	YES	YES	YES
	Catalogue			YES	YES	YES	YES
	Technical Compliance Clause by clause			YES	YES	YES	YES
	Performance statement Sextion IX			YES	YES	YES	YES
	Order copies /End User Certificate			YES	YES	YES	YES
	Model			"INSTRUMED"	SISCO I	MEDICON-MADE IN GERMANY	B MARTIN
	Bid validity			180 days	180 days	180 days	180 days
	IT			YES	YES	YES	YES
	Mandate Form			YES	YES	YES	YES
	Payment terms			YES	YES	YES	YES
	Delivery period GCC Clause 15			NO	YES	YES	YES
	warranty/GCC clause 15			36 months	36 months	36 months	36 months
	P&L a/c and BS			YES	YES	YES	YES
	Eligibility criteria 1600 kits each 100% supply of similar equipments			YES	YES	YES	YES
	Section VIII Bidder Information			Submitted	Submitted	Submitted	Submitted
7	Quality management system						
	ISO			YES	YES	YES	YES
	CE			NO	YES	YES	YES
	Recommendation OK/ NOTOK			NOT OK	OK	OK	OK
	Prepared by						
	Signature						
	Date						