

MPPHSL Tender-30/39 Amendment-1 in Technocommercial Evaluation

S.N.	Sch No.	Equipment Name	Quantity	Total Bids	Names of participated Bidder(s)	Name of manufacturer and Quoted Model	Commercial Qualification	Technical Qualification	Overall Qualification	Remark(Reason for non-qualification)	Reason of Amendment
1	5	BP apparatus stand model	20	1	Ashok Enterprises	Honsun-china Model	Non-responsive	Non-responsive	Non-responsive	not supplied as per qualification criterea	
				2	Drug depot	A&D Instru. India of A&D Japan	Responsive	Responsive	Responsive	not supplied as per qualification criterea	After Query Required proof of performance(End user certi.) found in the physical bid
17	26	Glucometer With strip	32	1	Drug Depot	USA-nipro-true result	Responsive	Responsive	Responsive	not supplied as per qualification criterea	After Query Required proof of performance(End user certi.) found in the physical bid
28	57	Ventilator ICU	2	1	COVIDIEN	e360T-USA-OK	Responsive	Responsive	Responsive		
				2	SCHILLER	Neumovent graphnet TS-argentina	Responsive	Responsive	Responsive	compressures CE not submitted	After Query Checked the desired documents and found responsive.
				3	MINDRAY	synovent E5 china	Responsive	Responsive	Responsive		
				4	Rohanika	V8800 oricare USA	Responsive	Non-responsive	Non-responsive	not fit for newborn of <3.5kg	
				5	Medion	V8800 oricare USA	Non-responsive	Non-responsive	Non-responsive	manu. Autho. Form is not proper and ins. is not fit for newborn	
29	58	Ventilator Pediatric	3	1	COVIDIEN	PB840-Ireland	Responsive	Responsive	Responsive		

MPPHSC Tender-30/39 Amendment-1 in Technocommercial Evaluation

S.N.	Sch No.	Equipment Name	Quantity	Total Bids	Names of participated Bidder(s)	Name of manufacturer and Quoted Model	Commercial Qualification	Technical Qualification	Overall Qualification	Remark(Reason for non-qualification)	Reason of Amendment
				2	SCHILLER	NEUMOVENT- ARGENTINA	Responsive	Responsive	Responsive	compressures CE not submitted	After Query Checked the desired documents and found responsive.
				3	Rohanika	fabian HFO Swiss	Non-responsive		Non-responsive	Not supplied as per qualification criteria	
				4	Medion	V8800 oricare USA	Non-responsive	Non-responsive	Non-responsive	manu. Autho. Form is not proper and ins. Is not fit for newborn,3,5	
33	67	Cold Light Source (LED)	4	1	Vijay Sales	Promis /PMA- LL-1000- Korea	Responsive	Responsive	Responsive	not supplied as per qualification criteria	After query checked the PO and found it was supplied as a spare part of Laproscope and was in Desired Quantity
				2	Drug Depot	LS-LED 100- Medicon- Germany	Responsive	Responsive	Responsive		