

Response to the queries raised by prospective bidders in the pre bid meeting, Dated 14 Sep.2015/Tender number 42 /Mannequin .

S.No.	Issue raised during pre bid	Remark's
1	Adult CPR mannequin should have a "clicker" which confirms correct compression depth	Remains unchanged, as specifications are as per GOI guidelines.
2	Adult CPR mannequin should have weight:3.9 kg & H x W x D:25in x 8.5 in x 13.5in	Remains unchanged, as specifications are as per GOI guidelines.
3	Delivery period should be extended to 90-120 days instead of 60 days. How much amount will be deducted if delivery is made after 60 days.	Remains unchanged. 20% amount will be deducted if delivery is made after 60 days.


CGM (TECHNICAL)