

		NON RESPONSIVE	PARTIALLY RESPONSIVE	RESPONSIVE	
	Tender no 43 Technical evaluation report for dialysis consumables				
	Dialysis Disposables –Specifications -The requirement may increase or decrease by 50% depending on the number of cases and cycles done (1.Quantity may increase or decrease by 50% 2. All items to be USFDA & CE, ISO 13485:2003; STERILE; ETO/GAMMA)	a) M/s Health care Medical Devices Pvt Ltd	b) M/s Hemant Surgical Industries	c) M/s Meditech Devices Private Ltd	Items to be retendered
DIACON 001	<p>a) 1.3/1.4 Adult Dialyser as per Specifications Dialyzers (multiple use): Hollow fiber size as given below polysulphone or Polyethersulfone Like of: Fresenius/BAIN(DORA)/DIALIFE/NIPRO/TORAY/BRAUN/GAMBRO/BAXTER/BIOLIGHT or equivalent</p> <ul style="list-style-type: none"> • Low flux high performance dialyser • Performance enhanced technology with high biocompatibility <ul style="list-style-type: none"> • Housed in medical grade polycarbonate <ul style="list-style-type: none"> • Open able ends for easy cleaning • Caps for blood inlet and outlet ports for safer storage <p>Caps for dialysate inlet and outlet for safer storage open able caps (Red & Blue)</p> <ul style="list-style-type: none"> • High quality sterilisation procedure using gamma radiation <p>Dialyzers should be meeting following standards: European CE,ISO13485:2003 STERILE ;sterilised Packs must be marked :GOMP Supply not for Sale</p>				Retender
DIACON 002	<p>b)1.5/1.6 Adult Dialyser as per Specifications Dialyzers (multiple use): Hollow fiber size as given below polysulphone or Polyethersulfone Like of: Fresenius/BAIN(DORA)/DIALIFE/NIPRO/TORAY/BRAUN/GAMBRO/BAXTER/BIOLIGHT or equivalent</p>				retender

	<ul style="list-style-type: none"> • Low flux high performance dialyser • Performance enhanced technology with high biocompatibility <ul style="list-style-type: none"> • Housed in medical grade polycarbonate <ul style="list-style-type: none"> • Open able ends for easy cleaning • Caps for blood inlet and outlet ports for safer storage <p>Caps for dialysate inlet and outlet for safer storage open able caps (Red & Blue)</p> <ul style="list-style-type: none"> • High quality sterilisation procedure using gamma radiation <p>Dialyzers should be meeting following standards: European CE,ISO13485:2003 STERILE ;sterilised Packs must be marked :GOMP Supply not for Sale</p>				
DIACON 003	<p>c) 1.7/1.8 Adult Dialyser as per Specifications Dialyzers (multiple use): Hollow fiber size as given below polysulphone or Polyethersulfone Like of: Fresenius/BAIN(DORA)/DIALIFE/NIPRO/TORAY/B BRAUN/GAMBRO/BAXTER/BIOLIGHT or equivalent</p> <ul style="list-style-type: none"> • Low flux high performance dialyser • Performance enhanced technology with high biocompatibility <ul style="list-style-type: none"> • Housed in medical grade polycarbonate <ul style="list-style-type: none"> • Open able ends for easy cleaning • Caps for blood inlet and outlet ports for safer storage <p>Caps for dialysate inlet and outlet for safer storage open able caps (Red & Blue)</p> <ul style="list-style-type: none"> • High quality sterilisation procedure using gamma radiation <p>Dialyzers should be meeting following standards: European CE,ISO13485:2003 STERILE ;sterilised Packs must be marked :GOMP Supply not for Sale</p>				retender

<p>DIACON 004</p>	<p>d)high flux Dialysers HS 80 or equivalent Adult dialyser Dialyzers (multiple use): Hollow fiber size as given below polysulphone or Polyethersulfone Like of: Fresenius/BAIN(DORA)/DIALIFE/NIPRO/TORAY/B BRAUN/GAMBRO/BAXTER/BIOLIGHT or equivalent</p> <ul style="list-style-type: none"> • Low flux high performance dialyser • Performance enhanced technology with high biocompatibility <ul style="list-style-type: none"> • Housed in medical grade polycarbonate • Open able ends for easy cleaning • Caps for blood inlet and outlet ports for safer storage <p>Caps for dialysate inlet and outlet for safer storage open able caps (Red & Blue)</p> <ul style="list-style-type: none"> • High quality sterilisation procedure using gamma radiation <p>Dialyzers should be meeting following standards: European CE,ISO13485:2003 STERILE ;sterilised Packs must be marked :GOMP Supply not for Sale</p>				<p>Retender</p>
<p>DIACON 005</p>	<p>A-V Blood Lines with AV Pressure Transducer Protector for all machines types and dialysers (Acceptable for fitting to all standard Dialyzers) with side tubing-(FOR Heparinization and AV Pressure monitoring) CE and ISO certificate essential with Medical grade PVC Like of :Fresenius/BAIN (DORA)/Dialife/NIPRO/B BRAUN/Browndone/Biolight or equivalent Colour-coded clamps and dialyser connectors for easy identification A-V Blood line tubing with AV Pressure Transducer Protector should be meeting following standards: European CE,ISO13485, Sterile, FDA Packs must be marked :GOMP Supply not for Sale</p>			<p>Responsive</p>	
<p>DIACON 006</p>	<p>a) Femoral catheters Single Lumen Size Adult(Set of 12 different sizes) Like of: Like of:</p>			<p>Responsive</p>	

	<p>BROWNDOVE/Medicath/Arrow/S.T. Stone/Medcomp/Dr. Surgical/Nephrotech/Bard/joline/B Braun/New tech or equivalent</p> <p>Packs must be marked :GOMP Supply not for Sale</p>				
DIACON 007	<p>b)Femoral catheters Single Lumen Size Pediatrics(Set of 12 different sizes) Like of: Like of: BROWNDOVE/Medicath/Arrow/S.T. Stone/Medcomp/Dr. Surgical/Nephrotech/Bard/joline/B Braun/New tech or equivalent Packs must be marked :GOMP Supply not for Sale</p>			Responsive	
DAICON 008	<p>Femoral catheters: double Lumen Kit curved Double lumen catheter set 12F (curved) adult</p>			Responsive	
DIACON 009	<p>• Adult - Double lumen catheter set 11.5 F - 12Fr, 13 cm. (curved) Like of: Like of: BROWNDOVE/Medicath/Arrow/S.T. Stone/Medcomp/Dr. Surgical/Nephrotech/Bard/joline/B Braun/New tech or equivalent Packs must be marked :GOMP Supply not for Sale</p>			Responsive	
DIACON 010	<p>• Paediatric - Double lumen catheter set 6.5 F – 8.5 Fr, 13 cm (curved) Like of: Like of: BROWNDOVE/Medicath/Arrow/S.T. Stone/Medcomp/Dr. Surgical/Nephrotech/Bard/joline/B Braun/New tech or equivalent Packs must be marked :GOMP Supply not for Sale</p>			Responsive	
DIACON 011	<p>• Adult - Double lumen catheter set 11.5 F - 12Fr, 13 cm. (straight) Like of: BROWNDOVE/Medicath/Arrow/S.T. Stone/Medcomp/Dr. Surgical/Nephrotech/Bard/joline/B Braun/New tech or equivalent Packs must be marked :GOMP Supply not for Sale</p>			Responsive	
DIACON 012	<p>• Paediatric - Double lumen catheter set 6.5 F – 8.5 Fr, 13 cm (straight) Like of: Like of: BROWNDOVE/Medicath/Arrow/S.T. Stone/Medcomp/Dr. Surgical/Nephrotech/Bard/joline/B Braun/New tech or equivalent Packs must be marked :GOMP Supply not for Sale</p>			Responsive	

DIACON 013	<p>Jugular Catheters :Double Lumen Kit curved Like of: BROWNDOVE/Medicath/Arrow/S.T. Stone/Medcomp/Dr. Surgical/Nephrotech/Bard/joline/B Braun/New tech or equivalent Standard</p> <p>gle pouch packing</p> <p>ibacterial hydrophobic air filter with Luer lock</p> <p>gged housing designed to withstand pneumatic pulsing</p> <p>ecting patients and equipment from cross contamination</p> <p>Standard Filter Membrane Housing: Medical grade semi-rigid PVC Should meet the following standards: European CE,ISO13485, Sterile, FDA Packs must be marked :GOMP Supply not for Sale</p>			Responsive	
DIACON 014	<p>Femoral Guide wires :Straight (0.325MM sizes) Should meet the following standards: European CE,ISO13485, Sterile, FDA Packs must be marked :GOMP Supply not for Sale</p>			Responsive	
DIACON 015	<p>Blood Vessel introducers Needles 16G , Sterilized Packs must be marked :GOMP Supply not for Sale</p>			Responsive	
DIACON 016	<p>Haemo Dialysis Fluid :For B carb Mode: Part A 10 Lit Can Packs must be marked :GOMP Supply not for Sale Deleted as already finalized Haemo Dialysis Fluid :For B carb Mode: Part B 500gm 2packs Packs must be marked :GOMP Supply not for Sale</p>		Responsive		
DIACON 017	<p>a)Citric Acid Granules/20% Sol .Citric Acid Anhydrous AR,ACS meets specification of IP, BP, USP, Ph. Eur Storage in tight containers away from heat and humidity (75oF and 55% RH) is recommended. Packs must be marked :GOMP Supply not for Sale</p>				Retender

DIACON 018	b)Sodium Hypochlorite Sol 10%(Approx. 4% available Chlorine) Packs must be marked :GOMP Supply not for Sale				Retender
DIACON 019	c)Formalin(CH ₂ O) not less than 30% Packs must be marked :GOMP Supply not for Sale				Retender
DIACON 020	d)Hydrogen Peroxide solution about 6% w/v H ₂ O ₂ (20 volumes) Small Containers: Small containers of pure aluminium of high density polyethylene NOTE: The containers shall be labelled as per the provision of the relevant Acts and Rule. Leakage: If spills and leaks occur, wash off all contaminated surface with plenty of water. For Waste Disposal: All relevant regulations of local and state authorities for prevention and control of pollution should be observed. (Hydrogen peroxide decomposes of its own giving water and oxygen the product universally accepted and not creating an additional pollution). As per Indian Standards: UDC 661.491 : 614.878 BIS 1994 BIS Certification Marking The product may also be marked with the Standard Mark. Packs must be marked :GOMP Supply not for Sale				Retender
DIACON 021	e)Acetic Acid 30% Glacial Corrosive flammable liquid-Properly labled with warnings as per Occupational Safety and Health Guidelines Storage: Storage area should be detached and should be well ventilated and should have acid resistant floors.Electrical fixtures in the area is vapour proof and any tolls if used should be non sparking type. Stored separately and handles under supervision. No Spills and leaks				Retender

	allowed. Packs must be marked :GOMP Supply not for Sale				
DIACON 022	f)Oxalic Acid Granules- PURIFIED PACKING, STORAGE AND MARKING Packing The material shall be securely packed in well-filled containers with minimum access to light and moisture. The containers shall be such as to preclude contamination of the contents with metals or other impurities. Storage The material shall be stored in a cool and dry place so as to avoid excessive exposure to heat. IS 5057:1997 Marking 53.1 Each container shall be legibly and indelibly marked with the following information: Name of the material including the words `Food Grade'; Type of the material whether `anhydrous'; Name and address of the manufacture; Net content when packed; Batch or code number; Directions for storage; Expiry/Best before date; and Any other requirements as specified under the Standards of Weights and Measures (Packaged Commodities) Rules, 1977J Prevention of Food Adulteration Rules, 1955. BIS Certification Marking The product may also be marked with the Standard Mark. Packs must be marked :GOMP Supply not for Sale				Retender
DIACON 023	Dialysis starting kit-Disposable a) Sterile Tray with Top (disposable), size not less than 30cm x 30cm. b) Sterile drape size not less than 45cmx 45cm – 1 No. c) Cotton ball – 6 Nos. d) Cotton Gauze piece (1"x 1") (dressing pad) – 6 nos e) Artery forceps – 1 No. (disposable) Packs must be marked :GOMP Supply not for Sale	Non responsive		Responsive	
DIACON 024	FISTULA 16 G Single sterilized ETO Single needle packing Like of RADIANT, New-Tech, Dr-Sure-Atlantic, Dr. Surgical, Nipro, Rein		Responsive	Responsive	
DIACON 025	FISTULA 17 G Single sterilized ETO Single needle packing Like of RADIANT, New-Tech, Dr-Sure-Atlantic, Dr. Surgical, Nipro, Rein		Responsive	Responsive	

DIACON 026	Torniquet with belt (good quality pairs)			Responsive	
DIACON 027	Neomycin and polymyxin B sulphate and bacterial Zinc Powder (Antibiotic) for external use only The four main ingredients in each gram Neomycin Suphate IP 3400 units, Polymyxin B Sulphate USP 5000 units and Becitracin Zinc IP 400 units				Retender
Note:	The requirement may increase or decrease by 50% depending on the number of cases and cycles done				
	The quantity mentioned is only the tentative requirement and may increase or decrease as per the decision of Ordering Authority and /or Tender Inviting Authority. The rates quoted should not vary with the quantum of the order or the destination. All sterilized products should be BIS/CE/USFDA approved and marked (ISI/CE)/approved by US FDA. All quoted products should conform to the requirement of relevant national/international quality certifications. All relevant valid approvals/certificates should be attached along with the bid.				
	Refer: the Manufacture, Storage and Import of Hazardous Chemical (Amendment) Rules, 2000.Ministry of Environment & Forests New Delhi, the 19th January, 2000 and as amended from time to time				
	For all above items manufacturer to Submit :				
a)	Warranty or Shelf life to be mentioned by bidder				
b)	Test certificate from Manufacturer through a third party accredited lab to be submitted Manufacturer Batch Report to be send				
c)	Quality Assurance Approval: Approval (USFDA or CE or BIS –as applicable) to be submitted				
d)	Market Standing certificate for last 3 years				
e)	Declaration of fall clause by bidder				
f)	The quantity may increase or decrease by 50% depending on number of centers successfully started				
	Note:Any queries to be entertained till 1 PM 28 th Oct 2015 after which Financial bid would be opened the next day.				

Chief General Manager-Technical, MPPHSCL