

**Madhya Pradesh Public Health Services Corporation Limited
(A Government of Madhya Pradesh Undertaking)**

1st Floor Oil Fed Building Bhopal

Website: www.health.mp.gov.in

PHONE: 0755-2578910

Ref. No. 44/MPPHSCL/2015/

Dated:- 03 Nov.2015

Corrigendum-2

With reference to Tender No. 44 and pre-bid resolution uploaded on 26th October 2015, following amendments are also being made to Schedule No. 3- Cuatery Machine:

Schedule No.	Item Name	Provision in the specification of Tender Document	Amendment
3	Cautery Machine/ Electro Surgical Unit	The maxi display should allow to check at a glance which program is currently Being used (Monopolar, bipolar, cutting or coagulation), the set output power, the selected type of current, the selected switching system (hand or footswitch) and the selected neutral (i.e. dispersive) electrode.	Relaced as: Display should allow to check at a glance which program is currently being used (Monopolar1, Monopolar2, Bipolar, Cutting or coagulation), the set output power, the selected type of current, the selected switching system (hand or footswitch) and the selected neutral (i.e. dispersive) electrode.
		Pre-bid amendment: Under Accessories clause: Patient Plate- 100 nos. disposable or 2 nos. Reusable plates; Vessel Sealing Clamp: 50 disposable clamps or 2 reusable clamps; All accessories should also have European CE or US FDA approved products.	Replaced as: Under Accessories clause: Patient Plate- 100 nos. disposable or 2 nos. Reusable silicon plates (dual type); Vessel Sealing Clamp: 10 disposable clamps or 2 reusable clamps; Laproscopic Vessel Sealing probe (with inbuilt cutting): Reusable 1 no. or Disposable: 2 nos.; Rest accessories remain same.
		Additional features:	Add: capable of sealing vessels up to 7 mm diameter
			Add: Automatic stoppage of output in case of malfunction with acoustic and visual signal
			Add: System for neutral plate safety by continuous monitoring of contact quality and connection.
			Add: The bidder should submit relevant and applicable IEC certificate along with European CE/ US FDA certificate.

In case of any inconsistency, above amendments shall prevail. Bidder should quote keeping above amendments and amendments made on 26th Oct 2015. Any change in bid date submission shall be notified separately.

Other terms and conditions of the tender document remains the same.


CGM (Technical)
MPPHSCL