

**Madhya Pradesh Public Health Services Corporation Limited  
(A Government of Madhya Pradesh Undertaking)**

Tilhan Sangh Bhawan Arera Hills Bhopal

Website: [www.health.mp.gov.in](http://www.health.mp.gov.in)

PHONE: 0755-2578911

**Corrigendum No.07**

Ref. No:-47/MPPHSCL/Medical College-Equip/2015

Dated:- 22-12-15

A N.I.T. was upload on website for Online Tender for the Annual Rate Contract of equipments/Instruments to be Supplied & Installed at Various Hospitals of Government of Madhya Pradesh for a Period of one Year ((List- 3) Equipment for 05 Clinical Departments of Medical College Associated Hospitals:-Anesthesia, Obstetrics & Gynecology, ENT, Physiotherapy and Casualty Departments) e-tender No. 47/MPPHSCL/ Medical College-Equip /2015 Dt. 14-10-15

Bidders are also informed that following amendment is being made:-

Schedule No.	Item No.	Item Name	As provided	As Amended
1	31	E.T.O. Sterilizer	<b>UNIQUE FEATURES</b> 1. Special Software where cycle parameters can be programmed as per actual requirement of the load. 2. "ENVIRO-PRO" system which acts as an additional area safety feature by sucking air away from operator. 3. Unique "TRI AERA" system, to enhance removal of residual EO from the products. 4. Unique "LEAKALERT" "sequence which performs a leak test automatically at the beginning of each cycle, and if leakage is found, it alerts the operator and aborts the cycle. 5. Fault and time out messages are displayed for easy diagnosis by operator and not only as Fault numbers. 6. Unique "ROTA-HEAT" system which enhances all round heating and lower electric charges. 7. Unique "AUTO-MAN" feature which allows Manual operation in case of emergency break down of Auto system. 8. Price economical with more features to price ratio. 9. Designed to work in tropical conditions and does not go in fault if the humidity or temperature is high.	Deleted unique features. Added: - Minimum capacity- 8 cu ft. Should have USFDA/ European CE, Should meet CSA / IEC standards,USOSHA

Other terms and conditions of the tender document remains the same.

  
FC CGM (Technical)  
MPPHSCL