

Techno-Commercial Evaluation of Tender No. 05/MPPHSL/Moruary/2014 dated 20.10.2014

Schedule No. 1: Two body Mortuary Cabinets, Qty: 102

Revised on
18.3.2015

Bid Submission date & time: 13.11.2014, 1700 Hrs

Bid opening date & time: 14.11.2014, 1300 Hrs

Items to be procured: Two Body Mortuary Cabinets for districts hospitals

Corrigendum no. 1 & 2 dated 3.11.2014 and 8.11.2014 respectively and Amendment no.1 (prebid meeting dated 30.11.2014) issued.

	Bid Clause No.	Descriptions	Esteem Industries Inc, Jharmajri, Baddi (HP) OK Confirmed	Rattan Sales Corporation, Janakpuri, New Delhi-58 OK	Arise Pharma, Bhopal manufactured by manufacturer Genist ,Mohali.	S M Scietific Instruments Pvt. Ltd, Delhi SM OK letter reply copy of (300+23)
Manufacturer Name			Esteem Industries Inc, Jharmajri, Baddi (HP)	Rattan Sales Corporation, Janakpuri, New Delhi-58	M/s Genist Technology Pvt. Ltd, Mohali India	S M Scietific Instruments Pvt. Ltd, Delhi
Quoted Model			ESTM-101	1210-SMC-002, Make: Spancers	GTM-001	SMI-157
Eligibility and Qualificati on Criteria	19.1	Earnest money of Rs.4,00,000/- (Rupees four lakh only)	Submitted	Submitted	Submitted	Submitted
	Section- IX (a)	a. The Bidder should be (i) a Manufacturer in India or (ii) Indian subsidiary of foreign manufacturer or (iii) authorized dealer / distributor of a foreign Manufacturer.	Bidder is Indian manufacturer.	Bidder is Indian manufacturer.	Bidder is dealer of M/s Genist Technology Pvt. Ltd, Mohali India.	Bidder is Indian manufacturer.
	Section- IX (b) ii	In case, bidder is an Indian Manufacturer or Indian Subsidiary of foreign manufacturer, it should have supplied and installed in last three years from the date of tender opening, at least 100% similar equipment meeting major parameters of technical specifications, anywhere in India and which is functioning satisfactory.	OK	OK	NA	OK
	Section- IX (b) ii	In case, bidder is an authorized dealer / distributor of a manufacturer, it should have supplied and installed in last three years from the date of tender opening, at least 50% similar equipment meeting major parameters of technical specifications, anywhere in India of the same manufacturer and which is functioning satisfactory.	NA	NA	Total 97 qty M Cabinet supply orders listed out in Performance Statement. Out of these, only 2 nos. supplied to end users and rest to dealers.	NA
Tech Specs	Sr. No.	Description of Technical specifications				
	I	Description of Function:				

	1.1	Mortuary Cabinet is meant for keeping dead body under cool condition to prevent decomposition of body. It should be equipped with hinged door and locking arrangement. It should have sufficient internal space for handling of dead bodies. It should be easy to clean and maintain.	Yes	Yes	Yes	Yes
	2	Technical Specifications				
	2.1	Should have two separate compartments	Yes	Yes	Yes as per compliance	Yes
	2.2	Should have corrosion free interior and exterior construction, preferably 304 SS Grade or better	Yes as per compliance	Yes as per compliance	Yes as per compliance	Yes as per compliance
	2.3	Temperature range should be 2 to 6 deg C	Yes as per compliance	Yes as per compliance	Yes as per compliance	Yes as per compliance
	2.4	Refrigeration System: Air cooled	Yes as per compliance- Fan to maintain uniform temperature.	Yes as per compliance	Yes as per compliance	Yes as per compliance
	2.5	Should have LED / LCD Temperature Display with audio-visual alarm	Yes, LCD Temp Controller.	Yes	Yes as per compliance	Yes
	2.6	Should have internal lighting arrangement	Yes, vapor proof indocent light.	Yes	Yes as per compliance	Yes
	2.7	Should have audio-visual alarm	Yes as per compliance	Yes	Yes as per compliance	Yes
	2.8	Should have simplified cadaver handling	Yes as per compliance- 3 piece carriage assembly	Yes	Yes as per compliance	Yes
	2.9	Exterior dimension approximately: Width : 1050mm to 1200mm Depth : 2400 mm Height : 1780 mm to 1800 mm				
			OK	OK	OK	OK
			OK	OK	OK	OK
			OK	OK	OK	OK
	2.10	Insulation Material: Rigid Polyurethane Foam (CFC Free) with density 40 Kg/m ³ or more	Yes as per compliance	Yes as per compliance	Yes as per compliance	Yes as per compliance
	2.11	One piece corrosion free stainless steel tray with tubular edge and handles.	Yes as per compliance	Yes as per compliance	Yes as per compliance	Yes as per compliance
	2.12	Should supply one Loading Trolley	Yes as per compliance- Hydrolic System loading trolley	Yes as per compliance	Yes as per compliance	Yes as per compliance
	2.13	Should have locking system and 3 keys	Yes as per compliance	Yes as per compliance	Yes as per compliance	Yes as per compliance
	2.14	Power supply: 220 Volt / 50 Hz.	Yes as per compliance	Yes as per compliance	Yes as per compliance	Yes as per compliance
	2.15	Should be supplied with suitable Stabilizer	Yes as per compliance-4 Kw automatic Stabilizer	Yes as per compliance	Yes as per compliance	Yes as per compliance
	2.16	Quoted model should be USFDA / European CE approved	Yes CE approved (Class-I) at P.11	Yes CE approved (Class-I) at P.12	Yes	Yes CE approved (Class-I) at P.12
	RESPONSIVE/NON RESPONSIVE		Responsive	Responsive	Responsive	Responsive

Corrigendum no. 1 & 2 dated 3.11.2014 and 8.11.2014 respectively and Amendment no.1 (prebid meeting dated 30.11.2014) issued.

	Bid Clause No.	Descriptions	Esteem Industries Inc, Jharmajri, Baddi (HP) OK	Rattan Sales Corporation, Janakpuri, New Delhi-58 OK	Arise Pharma, Bhopal	S M Scietific Instruments Pvt. Ltd, Delhi OK
Manufacturer Name			Self	Self	M/s Genist Technology Pvt. Ltd, Mohali India	Self
Quoted Model			ESTM-101	1210-SMC-004, Make: Spancers	GTM-001	SMI-157
Eligibility and Qualificati on Criteria	19.1	Earnest money of Rs.4,00,000/- (Rupees four lakh only)	BG No. ----- dated 8.11.2014, Union Bank of India for Rs. --- Lakh valid up to 8.6.2015.	BG No. 12871GPER002014A dated 11.11.2014, ----- for Rs.-- Lakh valid up to.....		BG No. 0505114BG0002711 dated 3.11.2014, for Rs 4 Lakh valid up to 2.7.2015 issued by SBI, Delhi.
	Section- IX (a)	a. The Bidder should be (i) a Manufacturer in India or (ii) Indian subsidiary of foreign manufacturer or (iii) authorized dealer / distributor of a foreign Manufacturer.	Bidder is Indian manufacturer.	Bidder is Indian manufacturer.	Bidder is dealer of M/s Genist Technology Pvt. Ltd, Mohali India.	Bidder is Indian manufacturer.
	Section- IX (b) ii	In case, bidder is an Indian Manufacturer or Indian Subsidiary of foreign manufacturer, it should have supplied and installed in last three years from the date of tender opening, at least 100% similar equipment meeting major parameters of technical specifications, anywhere in India and which is functioning satisfactory.	DHS, Maharashtra order dated 13.3.2012 for 98 nos. and another order dated 13th March 2014 for 60 nos. two body cabinets (P.28 to 66). Various end user certificated for only 15 nos. quantity.	Bidder has enclosed list of supply of 130 M abinets. On scrutiny of individual orders attached with the bid (P.53 to 112), it was found that the actual qualified order count comes to approx. 70 nos.	NA	16 nos. supply qty. of similar equipment mentioned in Performance Statement at P. 83. Meeting the criteria.
	Section- IX (b) ii	In case, bidder is an authorized dealer / distributor of a manufacturer, it should have supplied and installed in last three years from the date of tender opening, at least 50% similar equipment meeting major parameters of technical specifications, anywhere in India of the same manufacturer and which is functioning satisfactory.	NA	NA	Total 72 qty Mortuary Cabinet supply orders listed out in Performance Statement. Out of these, only 2 nos. supplied to end users and rest to dealers.	NA
Tech Specs	Sr. No.	Description of Technical specifications				
	I	Description of Function:				
	1.1	Mortuary Cabinet is meant for keeping dead body under cool condition to prevent decomposition of body It should be equipped with hinged door and locking arrangement. It should have sufficient internal space for handling of dead bodies. It should be easy to clean and maintain.	Yes	Yes	Yes as per compliance	Yes

	2	Technical Specifications			Yes as per compliance	Yes
	2.1	Should have four separate compartments	Yes	Yes	Yes as per compliance	Yes
	2.2	Should have corrosion free interior and exterior construction, preferably 304 SS Grade or better	Yes as per compliance	Yes as per compliance	Yes as per compliance	Yes
	2.3	Temperature range should be 2 to 6 deg C	Yes as per compliance	Yes as per compliance	Yes as per compliance	Yes as per compliance
	2.4	Refrigeration System: Air cooled	Yes as per compliance- Fan to maintain uniform temperature.	Yes as per compliance	Yes as per compliance	Yes as per compliance
	2.5	Should have LED / LCD Temperature Display with audio-visual alarm	Yes, LCD Temp Controller.	Yes	Yes as per compliance	Yes as per compliance
	2.6	Should have internal lighting arrangement	Yes, vapor proof incocent light.	Yes	Yes as per compliance	Yes as per compliance
	2.7	Should have audio-visual alarm	Yes as per compliance	Yes	Yes as per compliance	Yes as per compliance
	2.8	Should have simplified cadaver handling	Yes as per compliance- 3 piece carriage assembly	Yes	Yes as per compliance	Yes as per compliance
	2.9	Exterior dimension approximately:				
		Width : 2110	OK	OK	OK	OK
		Depth : 2400 mm	OK	OK	OK	OK
		Height : 1780 mm to 1800 mm	OK	OK	OK	OK
	2.10	Insulation Material: Rigid Polyurethane Foam (CFC Free) with density 40 Kg/m ³ or more	Yes as per compliance	Yes as per compliance	Yes as per compliance	Yes as per compliance
	2.11	One piece corrosion free stainless steel tray with tubular edge and handles.	Yes as per compliance	Yes as per compliance	Yes as per compliance	Yes as per compliance
	2.12	Should supply one Loading Trolley	Yes as per compliance	Yes as per compliance	Yes as per compliance	Yes as per compliance
	2.13	Should have locking system and 3 keys	Yes as per compliance	Yes as per compliance	Yes as per compliance	Yes as per compliance
	2.14	Power supply: 220 Volt / 50 Hz.	Yes as per compliance	Yes as per compliance	Yes as per compliance	Yes as per compliance
	2.15	Should be supplied with suitable Stabilizer.	Yes as per compliance-4 Kw automatic Stabilizer	Yes as per compliance	Yes as per compliance	Yes as per compliance
	2.16	Quoted model should be USFDA / European CE approved	Yes CE approved (Class-I) at P.11	Yes CE approved (Class-I) at P.12	Yes at P.25	Yes CE approved (Class-I) at P.12
		Remarks				
		RESPONSIVE/NON RESPONSIVE	RESPONSIVE	RESPONSIVE	RESPONSIVE	RESPONSIVE