

Madhya Pradesh Public Health Services Corporation Limited
(A Government of Madhya Pradesh Undertaking)

Tilhan Sangh Bhawan Arera Hills Bhopal

Website: www.mpphscl.in

PHONE: 0755-2578911

Corrigendum No. 05

Ref. No:-55/MPPHSCL/Medical College-Equip/2015

Dated: - 05 -02-16

Bidders are also informed that following amendment is being made:-

S.No.	Schedule No.	Item no.	Item name	As Provided	As Amended
1	1	73	Plasma Sterilizer		This item is deleted.
2	5	2	High Frequency Ventilator	Air Compressor not mentioned	1. The SPECIFICATION OF Schedule No.5 Item No.2- HIGH FREQUENCY VENTILATOR is revised & should be read as enclosed in Annexure 2.

3	3	10	ECHO machine	system should have multi array probe technology for phased array, linear array, curved array and multiplane TEE Transducer	System should have multi array probe technology and should be supplied with :- 1. Linear probe 4-12MHZ 2. Phased Array for pediatric cardiac application 3-9MHZ 3. Phased Array for Adult cardiac application 2-4MHZ The rates of TEE Probe should be quoted optionally in a seperate envelope physically.
4	3	16	3D Echo Cardiography and Color Doppler system	Should have various scanning methods: (a) Electronic Linear (b) Electronic Convex (c) Phased Array Sector	Amended as Follows :- Item Name Replaced as- Echo Cardiography and Color Doppler system Should have various scanning methods: 1 Linear probe 4-12MHZ 2. convex Array 2-6 Mhz 2. Phased Array for pediatric cardiac application 3-9MHZ 3. Phased Array for Adult cardiac application 2-4MHZ
5	1	66	Holmium laser 20 watt	Should supply common size fiber like 200-230 Micron, 300- 500 Micron & 600-800 Micron	Deleted :- "Green light LBO laser 120W for photoselective Vaporization of prostate along with side firebiber 50 no and continous flow laser cystoscope and UPS.
6	1	1	Radiofrequency_ cautery	Deleted- Primary frequency 4 Mhz or more	Added- Primary frequency 4 Mhz or more

7	1	67	Impedance based bipolar RFA for endoluminal ablation.	should be supplied with 25 units of single use applicators	Replaced as- Bidder should supply "Applicator – 2 boxes of 10 pieces".
8	3	22	Video colonoscope	Channel diameter : 3.8 mm or more	Replaced as- Channel diameter : 3.7 mm or more
9	3	24	Endoscopy ultrasound system (one complete unit with radial and linear probe)	Added- Dynamic High Definition Tissue Harmonic Imaging facility should be available	Deleted- Dynamic High Definition Tissue Harmonic Imaging facility should be available
				Added- Should be able to connect external probes.	Deleted- Should be able to connect external probes.
10	4	2	Video bronchoscope adult	4. 15" or more Medical grade High Resolution Monitor with atleast one DVI Output	Replaced as- 15" or more Medical grade High Resolution Monitor
				The system should also be capable for use with other rigid and flexible endoscopes with the help of additional camera head and should have integrated anti-moiré and anti-grid filters	Deleted- The system should also be capable for use with other rigid and flexible endoscopes with the help of additional camera head and should have integrated anti-moiré and anti-grid filters

11	8	1	3D-4D color Doppler high resolution USG machine	The Specification of Schedule No. 8 Item No.1- 3D-4D color Doppler high resolution USG machine is revised & should be read as enclosed in Annexure 2.
12	1	27	DSA machine with biplanner C arm with DSA & CT facility	Added- Turnkey work for DSA machine as enclosed in Annexure 2.