

Pre-bid query

Tender No.61/Pediatric surgery & Blood bank unit

S.N	Schedule No.	Item no.	Item name	Various Queries Received on below Tended Technical specifications	Amendments
1	BB2	17	GEL MICRO TYPING SYSTEM	<p>REAGENT PACK READER</p> <p>Company Should have Quality control reagent like cross matching grouping & antibody screening</p> <p>Single antigen test card & antisera like Rh System (C,c,E,e,e,C) Kell, Duffy, lewis, Lutheran MNS</p> <p>Special Card for DCT positive patients that can detect IgG, IgM, IgA, C3d,C3b</p> <p>Syphilis Antibody test</p> <p>Special reagent for acid elution method</p> <p>Incubator should be based on pelitear heating dedicated for AHG test & other immunohematological test</p> <p>Should be US FDA approved products</p>	<p>Deleted - Reagent pack reader</p> <p>Replaced as- Company should have Quality Control</p> <p>Replaced as- Optional- Single antigen test card & antisera like Rh System (C,c,E,e,e,C) Kell, Duffy, lewis, Lutheran MNS</p> <p>Replaced as - DCT Card for further work up of DCT Positive sample</p> <p>Deleted - Syphilis Antibody test</p> <p>Deleted - Special reagent for acid elution method</p> <p>Replaced as- Incubator should have 24 card capacity & should be compataible with offered technique</p> <p>Should be US FDA/European CE approved product</p>
2	PS18	6	Bair Hugger Warming Therapy Unit	<p>Seemingly by an oversight they have mentioned Bair Hugger is name of one of the companies making these type of devices which may please be removed from the tender and in place generic name like convection warming therapy unit may kindly be used. In this system specs if you could please specify the replacement time for filter to be around 1000 hrs it will be in the interest of the institution as faster filter replacement will result in to higher running cost otherwise.</p>	<p>Replaced as- Convection warming therapy unit.</p> <p>Added- The replacement time for filter to be 1000 hrs</p>
3	PS33	14	BINOCULAR LOUP	<p>In the specification for binocular loupes at 5th line they have mentioned zeiss anti-reflective coating. zeiss is also name of one of the companies making surgical magnification loupes. It is requested that the name zeiss may kindly be removed</p>	<p>Deleted- Zeiss</p>
4	PS9	3	Telescope for Pediatric Laparoscopy	<p>Repeation</p>	<p>No change</p>

Sharma
25/11/2016

मुख्य महाप्रबंधक (तकनीकी)
मध्यप्रदेश पब्लिक हेल्थ
सर्विसेस कार्पोरेशन लि., भोपाल